

SOUTH CAROLINA LAW ENFORCEMENT DIVISION

4400 BROAD RIVER ROAD
COLUMBIA, SOUTH CAROLINA 29210

POLICY #: 8.12.24	DATE: 09/01/93	REVISION DATE:
TITLE: IMPLIED CONSENT - DISPLAY AND PRINTING OF RUN STEPS AND RESULTS		PAGE 1 OF 1

RESPONSIBLE AUTHORITY: SPECIAL AGENT IN CHARGE OF
FORENSIC SERVICES LABORATORY

RELATED STANDARDS/STATUTES/REFERENCES: S.C. Code of Laws, Section 56-5-2950
S.C. Code of Laws, Section 50-21-114
S.C. Code of Laws, Section 55-1-100
S.C. Code of Laws, Section 56-1-2130
S.C. Code of Laws, Regulation 73-1

GENERAL PURPOSE: To set forth policies for the administration of implied consent breath alcohol tests.

POLICY: The Division will establish procedures for properly conducting implied consent BAC DataMaster breath alcohol tests.

SPECIFIC PROCEDURES: The test date will be printed at the top of the breath analysis section of the evidence ticket. Only the final test results will be displayed and printed for blank tests and subject tests. The convention of blood alcohol percent (grams ethanol per 100 milliliters of blood) will be used. This is a percent by weight. The forensically and medically accepted alveolar air to blood conversion factor of 2100 to 1 will be used in converting breath alcohol measurements to blood alcohol measurements. This ratio means that 1 milliliter of blood will contain approximately the same amount of alcohol as 2100 milliliters of alveolar air. This conversion factor is defined in Regulation 73-1, S.C. Code of Laws, 1976, as amended. The numerical results displayed and printed for the blank tests and subject test will be truncated to hundredths of a percent. The results of run sequence procedures (3) - (5), (8), (9), (11) - (12) will appear on the ticket. The basic layout of this section is (the numbers in parentheses are the run sequence procedures):

BREATH ANALYSIS ----- TEST DATE mm/dd/yy

BLANK TEST (3)	0.??%	hh:mm	
INTERNAL STANDARD (4)	VERIFIED	hh:mm	
0.10% SIMULATOR TEMP	34c	hh:mm	
0.10% SIMULATOR (5)	PASSED	hh:mm	
BLANK TEST (8)	0.??%	hh:mm	
SUBJECT SAMPLE (9)	0.??%	hh:mm	(or REFUSED)
BLANK TEST (11)	0.??%	hh:mm	
INTERNAL STANDARD (12)	VERIFIED	hh:mm	
REFUSAL SLIP #	????????		(if applicable)