A Nationally Accredited Agency

Certificate/Report of Evaluation

Manufacturer: Fluke
Model: 77
Description: Multimeter
Serial Number: 34753938
Technician: Stewart M. Robertson
Report Number: 1-34753938

Test Results: Passed
Date Checked: 09-17-2001
Due Date: 09-17-2002
Received: In tolerance
Returned: In tolerance

Reference Used

Manufacturer	*Model*	*Description*	*Cal Date*	*Due Date*
Fluke	45	Multimeter	06-13-2001	06-13-2002

Test Data

Test Points (BAC DataMaster #921012)	*Reference*	*Test Meter*	*Acceptable Range*
TP1 and TP5 (Sample Chamber Board)	2.40 V_{DC}	2.398 V_{DC}	+/- 0.3% (2.39 – 2.41 V_{DC})
TP0 and TP2 (CPU Board)	2.00 V_{DC}	1.997 V_{DC}	+/- 0.3% (1.99 - 2.01 V_{DC})
TP4 and bottom of R45 (Detector Board)	120.0 V_{DC}	119.7 V_{DC}	+/- 0.3% (119.6 – 120.4 V_{DC})

Comments: AS FOUND DATA IN TOLERANCE. NO ADJUSTMENT REQUIRED.