



CalNet®

Certificate of Calibration



Worldwide Support Center

1420 75th Street S.W.
Everett, WA 98203 USA
Phone: (888) 993-5853
Fax: (425) 446 6390

Manufacturer: FLUKE
Model: 45
Description: DUAL DISPLAY MULTIMETER
Asset Number: 10969
Serial Number: 6351007

Fluke certifies that the above listed instrument meets or exceeds all specifications as stated in the referenced procedure unless otherwise noted. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology (NIST), to NIST accepted intrinsic standards of measurement, or are derived by the ratio type of self-calibration techniques. This calibration complies with MIL-STD-45662A and ANSI/NCSL Z540.1-1994 (R2002).

This report may not be reproduced, except in full, unless permission for the publication of an approved abstract is obtained in writing from the calibration organization issuing this report.

Note: Any Test Uncertainty Ratio (TUR) that is less than four to one will appear under the "TUR" heading on the data record. If the TUR meets or exceeds four to one, the field is left blank.

CALIBRATION INFORMATION

Cal Date: 04-Sep-2003 Temperature 23°C Calibration Report Number: 827543-6351007:1062663202
Next Cal Due: 04-Sep-2004 Humidity 53 % Received Condition: IN TOLERANCE
Technician#: 16780 Returned Condition: IN TOLERANCE
Technician: Chris Bruner Pass Y

Remarks:

Calibration Procedure FLUKE 45: (1 YEAR) RS-232C CAL VER/ALT FLUKE 5700A Revision 1.2

Standards Used

Asset	Manufacturer	Model	Description	Cal Date	Due Date
9513	FLUKE	5725A	AMPLIFIER	16-May-2003	16-Nov-2003
9693	FLUKE	5700A	CALIBRATOR	16-May-2003	16-Nov-2003

Chris Bruner
Technician

Authorized Signature

Calibration Report Number: 827543-6351007:1062663202
Certificate/Report of Calibration 04-September-2003